GENUS UNI Complete and Innovative.

Options.

Two possible primary fixation.

Both tibial and femoral components are available in cemented or cementless version.



Genus UNI cementless femur.

Manufactured using the technology of powders, is characterized by three-dimensional Co-Por[®] monolithic surface HA coated.



Agnification of Ti-Por[®] Surface.

Genus UNI – cementless tibial component.

Manufactured using the technology of powders. The bone fixation is guaranteed by the three-dimensional Ti-Por[®] monolithic surface HAP coated.



Six femoral and tibial sizes.

Each tibial and femoral component has been studied to better fit with the internal or external compartment that it will replace.

Five tibial thickness for each size.

Both version "Metal-Back" and "All Poly" have 5 tibial thickness (9mm, 10mm, 11mm, 12mm, 14mm) to allow a millimetric optimization of the joint stability.

Genus UNI cemented femur.

It has been optimized with discharging compression of the cement in the anterior and posterior distal parts.



Genus UNI – Cemented tibial component.

There are anterior discharges for cementing in order to reduce the risk of leaving residual cement in the posterior part.

All Poly tibial component is also available





Adler Ortho srl Via dell'Innovazione 9 ADLER ORTHO UK 20032 Cormano, Italy Unit 16, M11 Tel +39 02 6154371 Business Link Fax +39 02 615437222 Personage Lane -

United Kingdom www.adlerortho.com Stansted Essex CM24 8GF Great Britain



B0022-00

| Femor | al Componer | nts | | | | |
|---------|-----------------|---------|--------------|--------------------------|---|--|
| SIZE | Ref. Codes Ce | emented | Ref.Codes Ce | Ref.Codes Cementless (*) | | |
| | RM/LL | LM/RL | RM/LL | LM/RL | | |
| 1 | 1182001 | 1182011 | 1115001 | 1115011 | | |
| 2 | 1182002 | 1182012 | 1115002 | 1115012 | | |
| 3 | 1182003 | 1182013 | 1115003 | 1115013 | | |
| 4 | 1182004 | 1182014 | 1115004 | 1115014 | | |
| 5 | 1182005 | 1182015 | 1115005 | 1115015 | | |
| 6 | 1182006 | 1182016 | 1115006 | 1115016 | | |
| Basi Ti | biali | | | | | |
| SIZE | Ref. Codes Ce | emented | Ref.Codes Ce | Ref.Codes Cementless (*) | | |
| | RM/LL | LM/RL | RM/LL | LM/RL | | |
| 1 | 1784001 | 1784011 | 1735001 | 1735011 | | |
| 2 | 1784002 | 1784012 | 1735002 | 1735012 | | |
| 3 | 1784003 | 1784013 | 1735003 | 1735013 | I | |
| 4 | 1784004 | 1784014 | 1735004 | 1735014 | | |
| 5 | 1784005 1784015 | | 1735005 | 1735015 | | |
| 6 | 1784006 | 1784016 | 1735006 | 1735016 | | |

(*) Ingrowth Surface Ti-Po+ HA Coating

| Inserts | | | | | | | | | | |
|---------|--------------|---------|---------------|---------|---------------|---------|---------------|---------|---------------|---------|
| SIZE | Ref.CodeS H9 | | Ref.CodeS H10 | | Ref.CodeS H11 | | Ref.CodeS H12 | | Ref.CodeS H14 | |
| | RM/LL | LM/RL | RM/LL | LM/RL | RM/LL | LM/RL | RM/LL | LM/RL | RM/LL | LM/RL |
| 1 | 1740901 | 1740911 | 1741001 | 1741011 | 1741101 | 1741111 | 1741201 | 1741211 | 1741401 | 1741411 |
| 2 | 1740902 | 1740912 | 1741002 | 1741012 | 1741102 | 1741112 | 1741202 | 1741212 | 1741402 | 1741412 |
| 3 | 1740903 | 1740913 | 1741003 | 1741013 | 1741103 | 1741113 | 1741203 | 1741213 | 1741403 | 1741413 |
| 4 | 1740904 | 1740914 | 1741004 | 1741014 | 1741104 | 1741114 | 1741204 | 1741214 | 1741404 | 1741414 |
| 5 | 1740905 | 1740915 | 1741005 | 1741015 | 1741105 | 1741115 | 1741205 | 1741215 | 1741405 | 1741415 |
| 6 | 1740906 | 1740916 | 1741006 | 1741016 | 1741106 | 1741116 | 1741206 | 1741216 | 1741406 | 1741416 |

All Poly tibial components

| SIZE | Ref.CodeS H9 | | Ref.CodeS H10 | | Ref.CodeS H11 | | Ref.CodeS H12 | | Ref.CodeS H14 | |
|------|--------------|---------|---------------|---------|---------------|---------|---------------|---------|---------------|---------|
| | RM/LL | LM/RL | RM/LL | LM/RL | RM/LL | LM/RL | RM/LL | LM/RL | RM/LL | LM/RL |
| 1 | 1850901 | 1850911 | 1851001 | 1851011 | 1851101 | 1851111 | 1851201 | 1851211 | 1851401 | 1851411 |
| 2 | 1850902 | 1850912 | 1851002 | 1851012 | 1851102 | 1851112 | 1851202 | 1851212 | 1851402 | 1851412 |
| 3 | 1850903 | 1850913 | 1851003 | 1851013 | 1851103 | 1851113 | 1851203 | 1851213 | 1851403 | 1851413 |
| 4 | 1850904 | 1850914 | 1851004 | 1851014 | 1851104 | 1851114 | 1851204 | 1851214 | 1851404 | 1851414 |
| 5 | 1850905 | 1850915 | 1851005 | 1851015 | 1851105 | 1851115 | 1851205 | 1851215 | 1851405 | 1851415 |
| 6 | 1850906 | 1850916 | 1851006 | 1851016 | 1851106 | 1851116 | 1851206 | 1851216 | 1851406 | 1851416 |

GENUS UNI Unicompartmental knee replacement



Tibial Screws

| CODE | DESCRIPTION | length(mm) |
|---------|------------------------|------------|
| 1710020 | Genus Uni Tibial Screw | 20 |
| 1710025 | Genus Uni Tibial Screw | 25 |
| 1710030 | Genus Uni Tibial Screw | 30 |
| 1710035 | Genus Uni Tibial Screw | 35 |
| 1710040 | Genus Uni Tibial Screw | 40 |
| 1710045 | Genus Uni Tibial Screw | 45 |
| 1710050 | Genus Uni Tibial Screw | 50 |

(*) Available on request

GENUS UNI Respect of the patients' anatomy.

- Main indications for unicompartmental knee arthroplasty
- Primary osteoarthritis of one compartment (medial or lateral).
- Good functionality of cruciate and collateral ligaments
- Not more than 15° of varus / valgus deformity and 10° of flexion deformitv
- Absence of any degenerative disease of progressive nature (eg Rheumatoid Arthritis)

Genus UNI has been designed according to the following principles:

- Respect for the morphology and anatomy of the patient. •
- Scrupulous respect for the ligamentous apparatus. •
- Minimum bone sacrifice •
- Logical, reproducible and minimally invasive surgical Technique. •

Instruments

The Instrument set allows the accurate reconstruction of the joint thanks to a special system of spacers.

Millimetric spacers can be added to the handle in order to check extension and flexion spaces and calibrate the bone resection depending on the cartilage wear of the patient.



Millimetric spacer to compensate cartilage wear in extension.

Handle for the control of spaces in flexion / extension

Millimetric spacer to determine the space in flexion.





In this way we can avoid overcorrection, responsible for the failure of many unicompartmental knee arthroplasty.





Component.













 Bone saving Balance of ligaments • Flexion Anatomical • Safety



Bone saving

6mm constant thickness of the implant. Better control of space in flexion and extension with minimal bone resection.

Balance of ligaments.

Rotation center at constant ra-

dius of flexion from 0 ° to 90 °

Flexion

The **3**° closure of the posterior condyle increases the range of motion and the stability of the system.



Anatomical

External distal apex blunted. Optimizes M/L positioning without risk of protrusion.

The prosthesis follows the anatomy of the medial and lateral condyles respecting the mechanical axis of the system.

Safety



Outer edge with softened radius. Minimizes the risk of damage on polyethylene and tip effect.

Tibial Component.





- Safety
- Stability
- Enhanced insert

Anatomical

The shape and dimensions of the tibial component have been studied to adapt to both the medial and lateral compartment maximizing the contact with the peripheral cortical bone.



Safety

Sistema di bloccaggio dell'inserto a coda di rondine per minimizzare i micromovimenti. Finitura a specchio per ridurre al minimo il rischio di generare debris di polietilene.

Anchoring pin for primary stabiliza-

tion of the tibial tray. Can be used

as a path for the insertion of a fixing

Anterior chamfer to avoid potential

screw (cementless option)

Enhanced Insert.

patellar conflicts.















Metal Back so to avoid potential conflicts between metal and soft parts.

Outer edge rounded to cover the

Flat articular surface to facilitate a linear contact with the condyle and better distribute the loading forces.











